


BGI INCORPORATED

 58 Guinan Street Waltham MA 02451
 Tel: 781-891-9380 Fax: 781-891-8151


CE Compliant Equipment

The following BGI Equipment is CE compliant for the intended purpose of the instrument due to the low internal voltages utilized (12 v dc). Compliance is immediately appertaining. They fall far below the CE Low Voltage standard of between 75 and 1,500 v dc. All motors and power supplies utilized are CE marked and fully compliant in their intended use.

- PQ100
- PQ200
- PQ167HA
- frmOmni
- deltaCal
- triCal
- AFC123
- Bag Filling Pump
- BGI400S
- Omni Personal Sampling Pump

The BGI Wright Dust Feeder (p/n WDF-II) has been tested to the following standards and is CE compliant accordingly.

- EN 61010-1:2001 (2nd Edition)
- IEC61010-1:2001 (2nd Edition)
- EN 61326: 1997/A1:1998/A2:2001/A3:2003

[Download a pdf copy of this page.](#)

**Ambient
Monitoring**
**Air Flow
Calibrators**
**BGI
Cyclone**
**Aerosol
Gen/Coll**
**Industrial
Hygiene**
**IAQ
Instruments**
**Draeger
Products**
**Filter
Media**
**Calendar
of Events**
Links
**Technical
Information**
**BGI
History**
**E-Mail
BGI**
Home